



Operating Manual

ABPM


1 Safety

2 Hardware

3 Software

4 Hygiene

Part 4: Hygiene and disinfection for custo screen 300/400/pediatric



Operating characteristics:

for custo screen 300/400/
pediatric with custo diagnostic

MHY 0003 – DK 1704
Version 001 – 07/08/2017

CE 0123

 **custo·med**
EXCELLENCE IN DIAGNOSTICS



© 2017 custo med GmbH

This Operating Manual may not be copied in its entirety or in part, duplicated in any form or by any means or translated into another language without the prior written consent of custo med GmbH.

The manufacturer reserves the right to change the information in this Operating Manual without prior notice. The current version can be downloaded from our website: www.customed.de.

CAUTION:

This Operating Manual is part of a modular system consisting of four parts. All four parts must be downloaded from the Internet or from a CD to ensure the Operating Manual is complete.



Operating Manual

ABPM

1 Safety

2 Hardware

3 Software

4 Hygiene

Part 4: Hygiene and disinfection for custo screen 300/400/pediatric

Table of contents

4.1	Important notes.....	4
4.2	Disinfection of the custo screen 300/400/pediatric recorder.....	4
4.3	Disinfection of the carrying case and the belt.....	4
4.4	Disinfection of the blood pressure cuff.....	4
4.5	Approved cleaning agents and disinfectants.....	5
	Cleaning overview at a glance.....	6

4.1 Important notes

Use only cleaning agents and disinfectants that have been approved by custo med. Unsuitable agents could damage the device.

The devices must never be immersed in liquid or cleaned with too much liquid. Cleaning agents and disinfectants must not be sprayed directly on or into the device. Make sure that no moisture gets inside the device (e.g. via interface contacts).

Contacts must not become soiled or damaged.

Clean and disinfect the devices after each change of patient. Make sure that the exterior of the devices is always aesthetic and clean.

The device must not be connected to a power source during cleaning and disinfection.

4.2 Disinfection of the custo screen 300/400/pediatric

- ▶ Reprocessing type: Wipe disinfection

4.3 Disinfection of the carrying case and the belt

- ▶ Reprocessing type: Washing with a disinfectant in a washing net

4.4 Disinfection of the blood pressure cuff

- ▶ Reprocessing type: Wipe disinfection
- ▶ Note: Observe the instructions on the instruction leaflet provided with the blood pressure cuff.
- ▶ **Cleaning and disinfection:** Wipe the cuff with a damp cloth. If necessary, remove the bladder and wash the cuff cover with soap or disinfectant solution. After disinfection, rinse the cuff under clear water and allow to air dry.

- **Sterilisation:** The cuff can be gas sterilised with ethylene oxide (EtO). After sterilisation, the parts that have been exposed to EtO must be ventilated. All the relevant regulations and safety standards must be strictly observed. Never autoclave.

The cuff tube, especially the BNC plug, must never be immersed in liquids.

4.5 Approved cleaning agents and disinfectants

Wipe disinfection:

Meliseptol® Wipes sensitive (B.Braun)

Meliseptol® Foam pure (B.Braun) – (use with a soft, lint-free cloth)

- Note: Please observe the manufacturer's instructions

Washing with a disinfectant:

Eltra 40® (ECOLAB)

Eltra 40® Extra (ECOLAB)

- Note: Please observe the manufacturer's instructions



Machine wash at 40°C in a washing net, (do not spin)



Do not treat with harsh chemicals or bleach



Do not iron



Do not tumble dry

Additional products that can be used for the disinfection of the blood pressure cuff:

Cidex, Sporicidin, Mikrozid, 70% isopropyl alcohol, ethanol 70%, buraton liquid

- Note: Please observe the manufacturer's instructions

-
- Note: The approved disinfectants can be replaced by products from other manufacturers provided they are equivalent in terms of disinfection and material compatibility. For more information, contact your partner for hygiene and disinfection.



Cleaning overview at a glance

Product	Wipe disinfection	Washing with a disinfectant	Miscellaneous
custo cardio 100/110/130/300	X		
custo cardio 200/400 and electrodes			S
custo screen 300/400/pediatric	X		
Blood pressure cuff			S
custo flash 500/510/510V/501/501L	X		
custo watch	X		
custo kybe	X		
custo guard	X		
custo belt electrode belt		X	
custo belt extender		X	
custo ECG cable guard 3 /cable adapter	X		
custo spiro mobile			S
Carrying cases, carrying belt, neck strap		X	

S Special reprocessing is necessary – for more information refer to the respective Operating Manual

Manufacturer's contact details:

custo med GmbH
 Maria-Merian-Straße 6
 85521 Ottobrunn
 Germany

Phone: +49 (0) 89 710 98 - 00
 Fax: +49 (0) 89 710 98 - 10
 E-mail: info@customed.de
 Internet: www.customed.de

